



**Date:** 4 June 2020



**Request for Quotation No. UNFPA/EGY/RFQ/20/009**

Dear Sir/Madam,


We hereby solicit your quotation for the supply of the items named below:






**NB: Sample for each quoted item required to be submitted along with the quotation no later than the deadline**


Item N°	Product Name	Product Description	Unit of Measure	Quantity
1	Surgical Mask, type I, for patients, non-sterile, disposable	Type I (EN 14683) surgical or medical mask. . Bacterial filtering efficiency (BFE): equal to or greater than 95%. . Differential pressure (breathability)/Breathing resistance: equal to or less than 29.4 Pa/cm2. . Splash resistance: None.	Box of 50 	11000
2	Surgical Mask, type I, for patients, non-sterile, disposable	. Fabric, non-woven with outer layer impervious liquid splash resistant material, e.g. polyethylene. . Comprised of 3 non-woven folded layers, shape completely covering nose, mouth and chin. . Clearly identifiable inner and outer surfaces. . Malleable nose strip, made of aluminum, allowing a snug fit. . With attached 2 x 2 tie-straps, allowing correct fixation and securing at the back of the head, or ear loops. . Size (indicative): 15-19 cm x 9-11 cm (l x w). Unfolded 175 x 175 mm.	Box 20-25 	11000

		<ul style="list-style-type: none"> <li>. Latex-free, glass fibre-free</li> <li>. Non-sterile</li> <li>. Single use, disposable.</li> </ul>		
3	Surgical respirator, high-filtration, FFP2/N95, with Valve disposable	<p>5ply N95 /Ffp2/Kn95 Face Mask Respirator with Filter/Valve, Safe, breathable, and Hypoallergenic; and high filtration, capacity. Elastic ear-loops with no pressure to the ears. Two pre-attached, strong elastic straps, fitting (i) around top of the head, (ii) around base of the head. Color: white. Non-sterile. Single use, disposable. Each mask bares clear identification of (i) protection provided FFP2/N95, (ii) which side to wear up (nose), (iii) manufacturer's name, and (iv) model reference</p>	<p>Each</p> 	150
4	Surgical respirator, high-filtration, FFP2/N95, mask, no valve, non-sterile, disposable	<p>Material: non-woven filter layer. Polypropylene, polyester, polyethylene, aluminum. Meets the requirements of FFP2 or N95 (FFP2 or N95 must be written on the respirator itself). Filtration level: &gt; 95 % for particles from 0.1 to 0.3 micron. Total inward leakage (TIL): &lt;10% (N95) or &lt;8% (FFP2). Penetration of the filtering</p>	<p>Each</p> 	150

		<p>material &lt; 6% (NaCl and paraffine at 95 l/min with particles of 0.6 µm).</p> <p>Air permeability: &gt; 2 mm H<sub>2</sub>O</p> <p>Meets the requirements of type IIR:</p> <p>Bacterial filtration efficiency (BFE) &gt; or = 98%.</p> <p>Differential pressure (breathability) &lt; 49 Pa.</p> <p>Splash resistance pressure &gt; or = 120 mm hg (tested in accordance with ASTM F1862 standard).</p> <p>Shape of the mask: duckbill (folded horizontal width-wise), or cup-shaped.</p> <p>Good breathability with design that does not collapse against the mouth.</p> <p>Without valve.</p> <p>Respirator mask fits all face shapes, without inspiration/expiration air-leakage.</p> <p>Upper edge has integrated easy malleable nose bridge strip reducing fogging of protective eye-wear.</p> <p>Size nose bridge strip: 4 x 90 mm (w x l) (+/-10%).</p> <p>Two pre-attached, strong elastic straps, fitting (i) around top of the head, (ii) around base of the head.</p> <p>Color: white.</p> <p>Non-sterile.</p> <p>Single use, disposable.</p> <p>Each mask bares clear identification of</p> <p>(i) protection provided FFP2/N95,</p> <p>(ii) which side to wear up (nose),</p> <p>(iii) manufacturer's name, and</p>		
--	--	--	--	--

		(iv) model reference		
5	Face Shield	<p>Durable full face length safety shield, fog-resistant. Encloses a wide area of the face ear-to-ear and forehead to chin.</p> <p>Can be worn with glasses or goggles.</p> <p>Made of clear plastic and provides good visibility to both the wearer and the patient.</p> <p>Material, shield part: clear polycarbonate, thickness approx. 1 mm.</p> <p>Made of robust material which can be cleaned and disinfected.</p> <p>Impermeable to liquids.</p> <p>Antistatic.</p> <p>Flexible.</p> <p>Size shield, from headband down approx.: 25 x 30 cm (w x h)</p> <p>Adjustable length headband, integrated with the shield.</p> <p>Adjustable band to attach firmly around the head and fit snugly against the forehead.</p> <p>Width headband, approx.: 3 cm</p> <p>Front part of the headband is foam padded (length approx. 25 cm)</p> <p>Shield is anti-fog treated/coated (preferred).</p> <p>Outside is coated to prevent glare from reflection.</p> <p>Non sterile.</p> <p>Reusable.</p>	<p>Each</p> 	120
6	Ethyl Alcohol	Ethyl Alcohol 70% Antiseptic, with pump	Bottle of 250 ml	11000

				
7	Ethyl Alcohol	Ethyl Alcohol 70% Antiseptic, with pump	Bottle of 500 ml 	11000
8	Ethyl Alcohol	Ethyl Alcohol 70% Antiseptic, with pump	Bottle of 1000 ml 	11000
9	Hand Disinfectant, Sterillium	Isopropyl Alcohol 70% Chlorhexidine gluconate 2% Glycerol, Liquid/Gel with Pump.	Bottle of 250 ml 	11000
10	Hand Disinfectant, Sterillium	Isopropyl Alcohol 70% Chlorhexidine gluconate 2% Glycerol, Liquid/Gel with Pump.	Bottle of 500 ml 	11000

11	Hand Disinfectant, Sterillium	Isopropyl Alcohol 70% Chlorhexidine gluconate 2% Glycerol, Liquid/Gel with Pump.	Bottle of 1000 ml	11000
				

The goods are to be delivered maximum in ( 7 ) Seven days upon issuing of PO. The quotation shall be valid at least for One month after the closing date.

If you are interested in submitting a quotation for these items, kindly fill in the attached Quotation Form and send by email to the address indicated below:

[egypt.tenders@unfpa.org](mailto:egypt.tenders@unfpa.org)

Please submit your quotation in Egyptian currency, if the offer is quoted differently from what is required, shall be based only on UN Operational Exchange Rate prevailing at the time of competition deadline.

Your earliest response to this query; as well as submitting the required mandatory samples for all items, would be highly appreciated, but not later than Wednesday **10 June 2020, at 15:00 Cairo Time**

Note: Current UNFPA supplier policies apply to this solicitation and can be found at: <http://www.unfpa.org/suppliers>.

Best regards,

**Contact Person: Gina Shoukry**

**Tel. No. +2 01006504944**

**Email: [egypt.tenders@unfpa.org](mailto:egypt.tenders@unfpa.org)**



United Nations Population Fund  
70A Nahda St., Saryat El- Maadi, Cairo.  
P.O. 11435 Egypt  
E mail: [egypt.tenders@unfpa.org](mailto:egypt.tenders@unfpa.org)  
Website: <http://www.unfpa.org>

### **Quotation Form**

**Name of Bidder:** \_\_\_\_\_

**Date of Bid:** \_\_\_\_\_

**Request for Quotation No:** **UNFPA/EGY/RFQ/20/009**

**Currency of Bid price:** **Egyptian Pounds**

**Delivery time** (*weeks from receipt of order till dispatch*): \_\_\_\_\_

**Expiration of Validity of Quotation** (*The quotation shall be valid for a period of at least One Month after the Closing date.*): \_\_\_\_\_

#### **Price Schedule:**

Item No.	Unit Price	Quantity	Transportation cost to UNFPA CO Premises,	Total in EGP	Delivery schedule (days)
1-Surgical Face Masks box of 50		<b>11000</b>			
2-Surgical Face Masks box of 20-25		<b>11000</b>			
3-N95 Face Masks		<b>150</b>			
4- N95 Face Mask with Valve		<b>150</b>			
5-Face Shield		<b>120</b>			
6-Ethyl Alcohol 250ml		<b>11000</b>			
7-Ethyl Alcohol 500ml		<b>11000</b>			
8-Ethyl Alcohol 1000ml		<b>11000</b>			
9-Hand Disinfectant, Sterillium liquid/Gel 250ml					
10-Hand Disinfectant, Sterillium liquid/Gel 500ml		<b>11000</b>			
11- Hand Disinfectant, Sterillium liquid/Gel 500ml		<b>11000</b>			

**In your offer, please include:**



United Nations Population Fund  
70A Nahda St., Saryat El- Maadi, Cairo.  
P.O. 11435 Egypt  
E mail: [egypt.tenders@unfpa.org](mailto:egypt.tenders@unfpa.org)  
Website: <http://www.unfpa.org>

1. Specific technical specifications of products offered
2. Quality standard of the products

*Vendor's Comments:*

I hereby certify that this company, which I am duly authorized to sign for, accepts the terms and conditions of UNFPA (<http://www.unfpa.org/resources/unfpa-general-conditions-contract> ) and we will abide by this quotation until it expires.

---

**Name and title**

---

**Date and Place**